

# Aortic Valve Endocarditis Post Ventricular Septal Defect Device Closure: Case Report

*Yasser Mubarak, MD., Minia University Correspondence:  
-Yasser Mubarak*

## **ABSTRACT:**

Ventricular septal defect (VSD) is a common congenital heart disease (CHD) in childhood, and its incidence is about 20% of CHD. Surgical closure or repair is safe and accepted procedure with minimal mortality. Transcatheter VSD closure offers excellent results. Coil system is developed for transcatheter VSD occlusion. Endocarditis is a rare but serious complication of device closure. Infective endocarditis (IE) represents a surgical challenge associated with perioperative mortality (1).

recommended as a standard surgical procedure (6).

The risk of infective endocarditis (IE) is increased. In the past, antibiotic IE prophylaxis was recommended in risk situations for transient bacteremia. Nowadays, prevention of IE focuses on prevention measures like good dental health. In addition, IE is a rare occurrence with decreasing mortality (7).

## **KEYWORDS:**

Ventricular septal defect, device closure, congenital heart disease, infective endocarditis, Aortic valve endocarditis

## **INTRODUCTION:**

Ventricular septal defect (VSD) is one of the most common form of congenital cardiac disease (CHD) in children especially perimembranous type (2). Surgical closure of VSD carries a minimal risk of morbidity and mortality, however transcatheter closure is less invasive, has a faster recovery and requires less hospital stay time (3). Percutaneous techniques have been developed in order to reduce the impact of such drawback of surgery (4).

## **CASE REPORT**

A 25-year-old male patient has history of previous device closure of perimembranous VSD during childhood. The transcatheter VSD closure was successful with no residual shunt, intraoperative or postoperative complications. He presented by a persistent fever of up to 40°C which failed to respond to one week course of antibiotics.

. Transcatheter VSD closure avoids extracorporeal circulatory support of surgical approach and has the benefit of shorter recovery time (5)

. The coil system is developed specifically for transcatheter occlusion of VSD. The incidence of endocarditis after device closure of VSD is extremely rare (3)

.The aortic valve is the most common site in IE, and its incidence is about 40% - 67%. Surgical intervention for aortic valve IE is about 60% - 72%. Aortic valve replacement (AVR); is cultures were taken, and all grew *Pseudomonas aeruginosa*. Echocardiography revealed vegetation of 20 mm × 5 mm attached to VSD device and aortic and moderately severe aortic insufficiency (AI). The VSD was completely closed by the coil with no residual shunt. The diagnosis of IE was based on the results of echocardiography and blood cultures, and made according to the Duke criteria. The patient was diagnosed with IE and started empirically on vancomycin. He showed no improvement and continued to develop intermittent fever; blood tests showed persistent leukocytosis, a high ESR, and high CRP. Also, he developed anemia, dyspnea, and signs of heart failure. Urgent surgical intervention underwent on cardiopulmonary bypass (CPB). Aortotomy and right atriotomy revealed a destroyed aortic valve (right coronary cusp) with vegetation and necrotic tissues around the coil. Before surgery, brain imaging is reasonable because ischemic/hemorrhagic stroke or mycotic aneurysm will delay surgical intervention and may require treatment. During surgery, care must be taken to avoid contamination of the field with infected tissue, and complete removal of foreign body is necessary before replacement with a valve. Infected leaflet and

The patient was admitted to the cardiology for additional investigation. The temperature was 39°C, heart rate was 140 bpm, respiratory rate was 30 bpm, and blood pressure was 100/60 mm Hg on admission. A new diastolic murmur (4/6) was detected in the aortic area with no organomegaly, no splinter hemorrhages or Roth's spots detected. Laboratory investigations showed an elevated erythrocyte sedimentation rate (ESR) of 140 mm/h, C-reactive protein (CRP) of 98 mg/dl, a

total leucocytic count (TLC) of 30,000/mm<sup>3</sup>. Three sets of blood

surrounding infected tissue were carefully inspected, entirely removed and sent to the laboratory for culture examination. The coil was extracted, *Fig.(1)* and the VSD was closed by Dacron patch using a continuous Prolene suture. Aortic valve replaced by size 23 bioprosthetic valve. Tricuspid valve tested and repaired. The bypass and cross clamp time were 37min. and 15 min, respectively. He required moderate doses of inotropes (adrenaline, dopamine, and dobutamine) and was continued on antibiotics. Cultures obtained from surgical specimens were sent.

## DISCUSSION

Transcatheter VSD closure is reported to be a safe and effective method, and has 94.8% a success rate. The early complications include heart block, device embolization, aortic insufficiency, tricuspid insufficiency, hemolysis, and residual shunts. However, Device closure is a foreign body; IE is considered a rare complication. Only a few cases with Amplatzer-device associated endocarditis have been reported in children after closure of atrial septal defects and two

In our case report, IE started at site of coil device and extended to aortic valve. Large vegetation, presence of foreign body, and developed of severe AI required urgent surgical intervention.

Surgical closure has been the standard method of VSD closure. However, although it is generally a safe procedure, it does have some potential risks: complete AVB in 1-5 % of the patients, significant residual VSD in 1-5 % with the necessity to re-operate in 2 %, and perioperative death in 0.5 %. Furthermore sternotomy, general anesthesia and cardiopulmonary bypass may initiate significant co-morbidities, such as infection, arrhythmias, and neurologic complications.

So, our case report reached old age with perimembranous VSD without any complications, and he and his family preferred transcatheter device closure to avoid any surgical co-morbidities.

AV repair for active endocarditis seems to lead to better survival compared with replacement, however, the use of large patches in combination with bicuspid anatomy results in increased risk of late failure (8).

We preferred replaced aortic valve with bioprosthetic valve to avoid long CPB time during trail of repair with lack our experience especially in critical sick patient.

cases were reported after VSD closure (3).

Pediatric cardiologists have been interested in closing VSDs interventionally for several years. Moderate success was achieved with initially large delivery systems and devices, where residual shunting, induction of regurgitation of the valves (tricuspid and aortic valve), as well as arrhythmia and hemorrhage were dangers. A transcatheter approach for the treatment of congenital heart diseases is appreciated by patients and their parents because of its lesser psychological impact (absence of a skin scar), shorter hospital time, less pain and discomfort, and avoidance of admission to an intensive care unit. Therefore, in the current era, transcatheter closure of perimembranous VSD provides a valuable alternative to surgery (5).

Transcatheter VSD device closure is safe and avoids complications of surgical closure, however its complications are serious and may need urgent surgical intervention.

**Abbreviations:** - VSD: ventricular septal defect, CHD: congenital heart disease, IE: infective endocarditis, AVR: aortic valve replacement, ESR: erythrocyte sedimentation rate, CRP: C- reactive protein, TLC: total leucocytic count, CPB: cardiopulmonary bypass, AI: aortic insufficiency.

**Acknowledgement:** We thank the collaborating cardiologists and ICU staff.

**Funding Statement:** The author received no funding.

**Conflicts of Interest:** no conflicts of interest to report regarding the present case report

## CONCLUSION



Figure (1):- Transcatheter VSD closure coil with infected tissue, aortic leaflets, and tip of central venous cannula extracted

## REFERENCES:

- 1- Suxuan Liua, Feng Chena, Xueyan Dinga , Zhenzhen Zhaob , Wen Kec , Yan Yand , et al. Comparison of results and economic analysis of surgical and transcatheter closure of perimembranous ventricular septal defect. *European Journal of Cardio-Thoracic Surgery* 42 (2012) e157–e162
- 2- Lei Wang, Sancheng Cao, Jun Li, Lifang Yang, Yin Liu, Jun Ren, et al. Transcatheter Closure of Congenital Perimembranous Ventricular Septal Defect in Children Using Symmetric Occluders: An 8-Year Multi-institutional Experience. *Ann Thorac Surg.* 2012; 94: 592 – 598.
- 3- El-Sisi AM, Menaissy YM, Bekheet SA. Infective endocarditis following coil occlusion of perimembranous ventricular septal defect with the Nit-Occlud (®) Le device. *Ann Pediatr Card.* 2016; 9: 59 - 61.
- 4- Mario Carminati, Gianfranco Butera, Massimo Chessa, Joseph De Giovanni, Gunter Fisher, Marc Gewillig, et al. Transcatheter closure of congenital ventricular septal defects: results of the European Registry. *European Heart Journal* (2007) 28, 2361–
- 5- Reyadh A. Elgindy, Hala M. Agha, Hala S. Hamza, Sherief H. Abdel Rahman
- 6- Khaled M. Saleh. TRANSCATHETER CLOSURE OF VENTRICULAR SEPTAL DEFECTS IN INFANTS AND CHILDREN. *Al-Azhar J Ped.* 2016; 19(2): 1633 - 48.
- 7- Duc Trung Nguyen , Francois Delahaye, Jean-Francois Obadia, Xavier Duval, Christine Selton-Suty, Jean-Pierre Carteaux, et al. Aortic valve replacement for active infective endocarditis: 5-year survival comparison of bioprostheses, homografts and mechanical prostheses. *Euro J Cardiothorac Surg.* 2010; 37: 1025 - 32.
- 8- Annina D., Daniel Q., Oliver K., Walter K. Immediate and Long-Term Results of Transcatheter Closure of Patent Ductus Arteriosus - Comparison of Two Decades before and after Change in Antibiotic Infective Endocarditis Prophylaxis Guidelines. *Congenital heart disease.* 2022; 17(2): 215 -230.
- 9- Katharina Mayer, Diana Aicher, Susanne Feldner, Takashi Kunihara and Hans- Joachim Schäfers. Repair versus replacement of the aortic valve in active infective endocarditis. *Euro J Cardiothorac Surg.* 2012; 42: 122–127.