



Endotracheal tube markings – Reverse order – A RARE manufacturing defect

Editor,

Endotracheal tubes have incremental markings from the tip to characterize the distance from the tip to the level of incisor as the depth of insertion. We here describe a case of reverse order of markings on endotracheal tube causing confusion to the exact depth of insertion.

In a 2 year old child, with a Cormack-Lehane grade I on laryngoscopy, a size 4.0 mm internal diameter Kimvent* Microcuff* Endotracheal tube (Kimberly-Clark Microcuff®PET; Kimberly Clark, Health Care, Atlanta, GA, USA) was passed under vision. Tube was positioned at 11 cm at incisors but it seemed to be reasonably inward on visible inspection of tube. On auscultation, it confirmed our diagnosis of endobronchial intubation. Subsequently the tube was withdrawn by 1 cm and to our astonishment we saw marking of 12 cm at incisor level. Then we closely observed the numbers on the tube and found they were in the reverse order (Fig. 1). Decision to change the tube was made as it was unusual to have reverse markings on the tube and depth of insertion could not be ascertained. The defective tube was reported to the manufacturer to rule out similar defective tubes in circulation posing risk to other patients.

Accurate depth of endotracheal tube placement is necessary, espe-

cially in paediatric population [1]. Successful tracheal intubation depends on a three stage process of selecting a tracheal tube of appropriate diameter, inserting it into the trachea to the correct depth, and securing it in the correct position [2]. The correct position of tracheal tubes should be confirmed immediately after insertion. The chest should be inspected for equal and bilateral chest expansion and auscultated for equal and bilateral air entry. The importance of intubating depth marking and the presence of four precision bands cannot be emphasized enough. All endotracheal tubes have markings from tip in centimeters in incremental numbers [3]. However, we report a case of endotracheal tube where markings from distal to proximal end were in declining numbers which led to confusion. Though it's just a limited manufacturing defect, but there can be issues if tube is just fixed by length and not reconfirmed on auscultation especially in emergency by paramedics. We emphasize the importance of checking the markings on tube prior to tube insertion in addition to cuff assembly check and making sure the orientation of markings furthermore.

Financial support

No financial support has been taken from any firm for the work.




Fig. 1. Reverse order of markings on Microcuff tube.

Peer review under responsibility of Egyptian Society of Anesthesiologists.

<https://doi.org/10.1016/j.egja.2018.03.003>

Received 22 February 2018; Received in revised form 15 March 2018; Accepted 16 March 2018

Available online 23 March 2018

1110-1849/ © 2018 Egyptian Society of Anesthesiologists. Production and hosting by Elsevier B. V. All rights reserved. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Conflict of interest

None.

References

- [1] De Caen A, Duff J, Coovadia AH, et al. Airway management. In: Nichols D, Ackerman A, Argent A, editors. *Rogers' textbook of pediatric intensive care*. 4th ed. Philadelphia: Lippincott Williams & Wilkins; 2008. p. 303–23.
- [2] Luten R. Accurate endotracheal tube placement in children: depth of insertion is part of a process. *Pediatr Crit Care Med* 2005;6:606–8.
- [3] Cook T. Airway management equipment. In: Davey A, Diba A, editors. *Ward's anaesthetic equipment*. 6th ed. Philadelphia: Elsevier Saunders; 2012. p. 139–205.

Neha Hasija*

VMMC & Safdarjung Hospital, New Delhi, India
E-mail address: hasijaneha@gmail.com

Subrahmanyam Maddirala

Department Anaesthesia, Rainbow Children's Hospital, Hyderabad, India
E-mail address: msubrah@gmail.com

Sunidhara Reddy

Rainbow Children's Hospital, Hyderabad, India
E-mail address: sunidhara@gmail.com

* Corresponding author at: A-77, First Floor, Malviya Nagar, New Delhi 110017, India.